
Midmark M11 - v1.0.0
Total Cycles: 5142

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2024 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 174.3 F 0.1
4:00 189.9 F 0.9
6:00 204.7 F 3.9
8:00 217.6 F 7.8
10:00 234.3 F 12.3
12:00 233.1 F 8.5
14:00 244.0 F 11.9
16:00 243.5 F 12.3
18:00 242.3 F 11.4
20:00 255.2 F 18.0
22:00 265.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.2 F 28.9
2:00 272.6 F 29.0
3:00 272.1 F 28.6
4:00 271.6 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 236.2 F 0.0
10:00 224.7 F 0.0
15:00 215.5 F 0.0
20:00 215.6 F 0.0
25:00 210.3 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:54
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:02:54
CYCLE COMPLETE

11/29/2024 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____