

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2538

Name: RITTER2  
Sterilizer ID: V2313228  
06/20/2023 11:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.7 F 0.0  
2:00 193.1 F 3.7  
4:00 206.8 F 9.2  
6:00 222.8 F 8.0  
8:00 233.1 F 10.9  
10:00 234.0 F 8.7  
12:00 239.7 F 10.1  
14:00 248.1 F 14.6  
16:00 260.5 F 21.5  
18:00 271.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.7 F 29.6  
2:00 271.8 F 28.8  
3:00 271.9 F 29.4  
4:00 272.7 F 29.4  
  
Min 271.3 F 28.6  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 214.1 F 0.0  
10:00 226.1 F 0.0  
15:00 230.6 F 0.0  
20:00 225.8 F 0.0  
25:00 225.5 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 2:11  
HEATING: 18:41  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:04  
CYCLE COMPLETE

06/20/2023 12:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_