
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001262
14:53 19/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 3:02
STERILIZING 4:07
136.2 C 220kPa 4:07
136.4 C 224kPa 4:37
136.4 C 221kPa 5:07
136.4 C 224kPa 5:37
136.4 C 222kPa 6:07
136.1 C 222kPa 6:37
136.4 C 225kPa 7:07
136.4 C 225kPa 7:37
136.4 C 225kPa 7:37
VENTING 7:37
AIR DRYING 8:28
CYCLE COMPLETE 23:28

Digital Signature #
8037248252F99715
