
Midmark M11 - v1.0.5
Total Cycles: 2135

Name: RITTER2
Sterilizer ID: V2313228
03/09/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.5 F 0.0
2:00 158.2 F 1.5
4:00 180.9 F 5.0
6:00 200.7 F 9.7
8:00 219.1 F 6.3
10:00 223.2 F 9.0
12:00 235.7 F 10.0
14:00 240.3 F 11.7
16:00 239.2 F 10.4
18:00 253.0 F 17.2
20:00 264.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.2 F 29.2
2:00 272.4 F 29.7
3:00 271.5 F 28.9
4:00 272.3 F 29.6

Min 271.3 F 28.9
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 226.4 F 0.0
10:00 220.5 F 0.0
15:00 221.1 F 0.0
20:00 215.9 F 0.0
25:00 225.9 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

03/09/2023 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____