

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1579

Name: RITTER 1  
Sterilizer ID: V2178198  
02/01/2022 09:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 109.2 F 0.0  
2:00 117.1 F 0.2  
4:00 131.8 F 2.3  
6:00 155.7 F 5.3  
8:00 179.4 F 9.4  
10:00 206.5 F 8.4  
12:00 218.9 F 6.6  
14:00 230.8 F 11.0  
16:00 230.7 F 6.9  
18:00 243.8 F 12.7  
20:00 241.2 F 10.7  
22:00 253.5 F 16.9  
24:00 264.6 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.7 F 28.7  
2:00 272.4 F 28.6  
3:00 271.6 F 28.5  
4:00 273.2 F 29.3

Min 271.0 F 27.9  
Max 273.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 230.7 F 0.0  
10:00 231.2 F 0.0  
15:00 227.3 F 0.0  
20:00 228.2 F 0.0  
25:00 228.5 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 25:60  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:04:05  
CYCLE COMPLETE

02/01/2022 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_