
Midmark M11 - v1.0.5
Total Cycles: 597

Name: RITTER2
Sterilizer ID: V2313228
12/01/2021 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.0 F 0.0
2:00 167.5 F 3.2
4:00 185.4 F 7.9
6:00 209.7 F 12.5
8:00 221.8 F 8.1
10:00 235.7 F 12.5
12:00 239.2 F 11.5
14:00 239.1 F 10.5
16:00 252.8 F 17.4
18:00 264.8 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.1 F 29.2
2:00 272.5 F 29.5
3:00 272.8 F 30.0
4:00 272.1 F 29.6

Min 271.8 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 236.4 F 0.0
10:00 230.1 F 0.0
15:00 221.1 F 0.0
20:00 218.3 F 0.0
25:00 217.3 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

12/01/2021 04:39 PM
ACCEPTED: X
REJECTED:
Operator: U02