

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000820  
10:59 13/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:50  
PRESSURIZING 3:56  
STERILIZING 5:09  
136.4 C 231kPa 5:09  
136.1 C 230kPa 5:39  
136.4 C 231kPa 6:09  
136.2 C 231kPa 6:39  
136.4 C 231kPa 7:09  
136.2 C 229kPa 7:39  
136.4 C 232kPa 8:09  
136.3 C 231kPa 8:39  
136.3 C 231kPa 8:39  
VENTING 8:39  
AIR DRYING 9:32  
CYCLE COMPLETE 24:32  
-----

Digital Signature #  
345E23BDD392D4CF  
-----