
Midmark M11 - v1.0.0
Total Cycles: 3519

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2023 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.7 F 0.0
2:00 168.6 F 1.3
4:00 188.8 F 4.8
6:00 209.6 F 9.5
8:00 223.7 F 7.3
10:00 231.6 F 9.5
12:00 236.6 F 9.8
14:00 243.1 F 12.3
16:00 241.5 F 11.0
18:00 254.7 F 17.8
20:00 266.6 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.8 F 29.0
4:00 271.8 F 28.4

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 211.1 F 0.0
10:00 219.3 F 0.0
15:00 220.5 F 0.0
20:00 221.0 F 0.0
25:00 219.6 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

07/12/2023 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____