

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000876  
9:51 20/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:40  
STERILIZING 3:53  
136.0 C 228kPa 3:53  
136.2 C 231kPa 4:23  
135.9 C 229kPa 4:53  
136.4 C 232kPa 5:23  
136.4 C 232kPa 5:53  
136.5 C 233kPa 6:23  
136.4 C 233kPa 6:53  
136.2 C 230kPa 7:23  
136.2 C 230kPa 7:23  
VENTING 7:23  
AIR DRYING 8:15  
CYCLE COMPLETE 21:45  
-----

Digital Signature #  
4BB3DDFAF6CF2DB8  
-----