
Midmark M11 - v1.0.0
Total Cycles: 3030

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2023 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.0 F 0.0

C103: HEATUP MODE

** ITEMS NOT STERILE **

03/24/2023 08:12 AM