
Midmark M11 - v1.0.0
Total Cycles: 2757

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2023 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 159.8 F 0.0
4:00 173.5 F 1.3
6:00 194.9 F 5.7
8:00 215.8 F 10.4
10:00 224.8 F 7.0
12:00 237.4 F 10.6
14:00 233.3 F 7.6
16:00 243.9 F 12.0
18:00 245.5 F 12.8
20:00 258.3 F 19.5
22:00 269.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.7 F 28.8
2:00 271.7 F 28.1
3:00 271.6 F 28.4
4:00 272.6 F 28.8

Min 271.1 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 234.6 F 0.0
10:00 230.7 F 0.0
15:00 229.3 F 0.0
20:00 231.6 F 0.0
25:00 231.9 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 23:10
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

01/10/2023 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____