
Midmark M11 - v1.0.5
Total Cycles: 2289

Name: RITTER2
Sterilizer ID: V2313228
04/24/2023 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 165.2 F 2.6
4:00 167.1 F 6.0
6:00 191.8 F 10.7
8:00 212.7 F 4.8
10:00 229.0 F 10.2
12:00 231.0 F 7.8
14:00 243.4 F 13.5
16:00 243.4 F 12.4
18:00 256.4 F 19.3
20:00 267.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.8
2:00 271.9 F 29.1
3:00 271.5 F 29.3
4:00 272.6 F 29.8

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 208.9 F 0.0
10:00 204.9 F 0.0
15:00 207.8 F 0.0
20:00 218.0 F 0.0
25:00 218.1 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

04/24/2023 10:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____