

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000852  
15:23 07/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:55  
STERILIZING 4:05  
136.0 C 229kPa 4:05  
136.4 C 233kPa 4:35  
136.4 C 233kPa 5:05  
136.4 C 233kPa 5:35  
136.3 C 232kPa 6:05  
136.2 C 231kPa 6:35  
136.4 C 233kPa 7:05  
136.3 C 232kPa 7:35  
136.3 C 232kPa 7:35  
VENTING 7:36  
AIR DRYING 8:27  
CYCLE COMPLETE 23:27

-----  
Digital Signature #  
D03FBB7B35D1EE0A  
-----