

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002253  
7:38 20/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:02  
STERILIZING 4:03  
136.4 C 222kPa 4:03  
136.4 C 223kPa 4:33  
136.0 C 220kPa 5:03  
136.4 C 222kPa 5:33  
136.4 C 220kPa 6:03  
136.3 C 220kPa 6:33  
136.4 C 223kPa 7:03  
136.0 C 221kPa 7:33  
136.0 C 221kPa 7:33  
VENTING 7:33  
AIR DRYING 8:23  
CYCLE COMPLETE 8:28

-----  
Digital Signature #  
B61F0B5246265CA4  
-----