
Midmark M11 - v1.0.5
Total Cycles: 664

Name: RITTER2
Sterilizer ID: V2313228
12/29/2021 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 192.1 F 4.0
4:00 208.2 F 9.8
6:00 223.0 F 7.2
8:00 234.3 F 11.6
10:00 236.2 F 9.8
12:00 236.9 F 9.1
14:00 250.4 F 16.0
16:00 262.4 F 22.9
18:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.3 F 29.3
2:00 272.5 F 29.9
3:00 271.5 F 29.0
4:00 272.5 F 29.6

Min 271.5 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 217.1 F 0.0
10:00 223.1 F 0.0
15:00 221.1 F 0.0
20:00 221.1 F 0.0
25:00 217.4 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:15
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:43
CYCLE COMPLETE

12/29/2021 01:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____