
Midmark M11 - v1.0.5
Total Cycles: 3630

Name: RITTER2
Sterilizer ID: V2313228
05/21/2024 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.2 F 0.0
2:00 173.1 F 2.4
4:00 186.4 F 6.9
6:00 207.7 F 12.2
8:00 221.7 F 7.6
10:00 234.7 F 13.2
12:00 238.6 F 10.9
14:00 237.1 F 9.6
16:00 251.5 F 16.6
18:00 263.5 F 23.5
20:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.2 F 29.5
2:00 271.3 F 28.8
3:00 272.5 F 29.7
4:00 271.9 F 29.0

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 233.7 F 0.0
10:00 197.8 F 0.0
15:00 193.2 F 0.0
20:00 188.7 F 0.0
25:00 186.3 F 0.0
30:00 191.0 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

05/21/2024 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____