

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1643

Name: RITTER2  
Sterilizer ID: V2313228  
10/24/2022 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
2:00 189.0 F 3.6  
4:00 201.3 F 8.3  
6:00 214.3 F 11.3  
8:00 224.9 F 9.2  
10:00 233.6 F 9.1  
12:00 242.3 F 12.8  
14:00 242.0 F 11.8  
16:00 255.4 F 18.7  
18:00 266.9 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.5 F 29.8  
2:00 272.0 F 29.2  
3:00 271.0 F 29.0  
4:00 272.6 F 29.9  
  
Min 271.0 F 28.8  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 213.8 F 0.0  
10:00 244.8 F 0.0  
15:00 253.0 F 0.0  
20:00 240.8 F 0.0  
25:00 226.2 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

10/24/2022 10:53 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_