
Midmark M11 - v1.0.0
Total Cycles: 3846

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2023 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.1 F 0.0
2:00 183.8 F 0.1
4:00 191.2 F 3.1
6:00 209.2 F 7.8
8:00 225.5 F 12.9
10:00 231.2 F 8.5
12:00 241.2 F 11.3
14:00 242.9 F 11.9
16:00 240.8 F 10.6
18:00 254.2 F 17.2
20:00 265.6 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.4 F 28.5
2:00 271.5 F 28.1
3:00 272.5 F 28.5
4:00 271.5 F 27.9

Min 271.4 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 201.5 F 0.0
10:00 200.1 F 0.0
15:00 200.3 F 0.0
20:00 216.1 F 0.0
25:00 227.6 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

10/02/2023 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____