

-----  
Midmark M11 - v1.0.5  
Total Cycles: 45

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2021 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.8 F 0.0  
2:00 124.3 F 1.5  
4:00 139.5 F 4.6  
6:00 169.1 F 8.6  
8:00 193.0 F 13.2  
10:00 206.5 F 7.0  
12:00 222.2 F 11.8  
14:00 229.4 F 9.3  
16:00 236.5 F 10.0  
18:00 244.4 F 14.3  
20:00 257.8 F 20.7  
22:00 267.9 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 30.2  
1:00 272.2 F 30.8  
2:00 271.9 F 30.2  
3:00 272.4 F 30.9  
4:00 272.1 F 30.3

Min 271.5 F 30.1  
Max 273.0 F 31.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
5:00 239.1 F 0.0  
10:00 222.5 F 0.0  
15:00 207.4 F 0.0  
20:00 195.3 F 0.0  
25:00 192.4 F 0.0  
30:00 192.2 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 23:25  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:01:45  
CYCLE COMPLETE

05/26/2021 12:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_