
Midmark M11 - v1.0.5
Total Cycles: 2405

Name: RITTER2
Sterilizer ID: V2313228
05/24/2023 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 183.0 F 2.4
4:00 196.0 F 7.2
6:00 216.3 F 12.8
8:00 226.3 F 8.6
10:00 237.8 F 10.7
12:00 242.0 F 12.3
14:00 241.5 F 11.3
16:00 254.9 F 18.2
18:00 266.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.4
2:00 271.5 F 29.0
3:00 272.6 F 29.4
4:00 271.5 F 28.8

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 230.2 F 0.0
10:00 244.3 F 0.0
15:00 238.1 F 0.0
20:00 226.9 F 0.0
25:00 222.5 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

05/24/2023 09:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____