
Midmark M11 - v1.0.5
Total Cycles: 2314

Name: RITTER2
Sterilizer ID: V2313228
04/28/2023 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.9 F 0.0
2:00 124.7 F 0.0
4:00 144.8 F 2.2
6:00 179.9 F 6.0
8:00 200.5 F 10.7
10:00 213.7 F 4.2
12:00 226.2 F 9.0
14:00 237.3 F 12.5
16:00 238.1 F 10.7
18:00 236.3 F 8.7
20:00 249.2 F 15.2
22:00 260.3 F 21.5
24:00 270.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.4 F 28.7
3:00 272.5 F 29.5
4:00 271.6 F 28.7

Min 271.3 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 211.4 F 0.0
10:00 229.3 F 0.0
15:00 227.7 F 0.0
20:00 226.9 F 0.0
25:00 228.7 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

04/28/2023 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____