

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2576

Name: RITTER2  
Sterilizer ID: V2313228  
06/27/2023 02:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.1 F 0.0  
2:00 138.9 F 0.0  
4:00 164.7 F 1.0  
6:00 188.1 F 5.1  
8:00 219.6 F 9.5  
10:00 228.0 F 10.1  
12:00 230.6 F 8.9  
14:00 239.5 F 11.9  
16:00 240.2 F 11.5  
18:00 239.1 F 10.3  
20:00 252.1 F 16.7  
22:00 263.2 F 23.3  
24:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.8  
1:00 272.0 F 29.3  
2:00 272.5 F 29.3  
3:00 271.5 F 28.7  
4:00 272.3 F 29.6

Min 271.3 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 231.2 F 0.0  
10:00 235.4 F 0.0  
15:00 223.4 F 0.0  
20:00 223.7 F 0.0  
25:00 230.4 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 24:09  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:02:28  
CYCLE COMPLETE

06/27/2023 03:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_