
Midmark M11 - v1.0.0
Total Cycles: 1300

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2021 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.5 F 0.0
2:00 94.9 F 0.0
4:00 106.4 F 0.0
6:00 140.0 F 0.0
8:00 184.1 F 0.1
10:00 205.4 F 0.2
12:00 217.2 F 2.2
14:00 224.4 F 4.4
16:00 231.2 F 6.7
18:00 238.8 F 9.7
20:00 247.1 F 12.8
22:00 242.5 F 11.3
24:00 238.1 F 9.3
26:00 240.1 F 9.8
28:00 248.8 F 14.4
30:00 259.9 F 20.4
32:00 269.8 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.9
1:00 272.4 F 28.5
2:00 272.3 F 28.2
3:00 271.9 F 28.2
4:00 272.7 F 28.5

Min 271.5 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 235.2 F 0.0
10:00 226.1 F 0.0
15:00 220.6 F 0.0
20:00 218.0 F 0.0
25:00 201.6 F 0.0
30:00 188.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:53
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:11:03
CYCLE COMPLETE

10/26/2021 10:32 AM
ACCEPTED: X
REJECTED:
Operator: U01