
Midmark M11 - v1.0.0
Total Cycles: 989

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2021 04:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 178.8 F 3.4
4:00 194.3 F 7.4
6:00 215.3 F 13.2
8:00 228.2 F 8.0
10:00 242.0 F 13.6
12:00 240.3 F 11.0
14:00 237.4 F 9.2
16:00 252.1 F 16.3
18:00 264.3 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.7
2:00 272.0 F 28.7
3:00 272.0 F 28.4
4:00 272.5 F 29.0

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 223.8 F 0.0
10:00 222.3 F 0.0
15:00 224.8 F 0.0
20:00 223.1 F 0.0
25:00 222.5 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

07/08/2021 05:26 PM
ACCEPTED: X
REJECTED:
Operator: U02