

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000989  
8:18 14/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:10  
STERILIZING 3:15  
136.0 C 228kPa 3:15  
136.3 C 229kPa 3:45  
136.2 C 231kPa 4:15  
136.2 C 229kPa 4:45  
136.4 C 230kPa 5:15  
136.4 C 233kPa 5:45  
136.4 C 229kPa 6:15  
136.2 C 231kPa 6:45  
136.2 C 231kPa 6:45  
VENTING 6:45  
AIR DRYING 7:37  
CYCLE COMPLETE 10:53  
-----

Digital Signature #  
59EA3103C9F8B200  
-----