
Midmark M11 - v1.0.0
Total Cycles: 1586

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2022 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 78.9 F 0.0
4:00 89.2 F 0.0
6:00 118.4 F 1.2
8:00 145.2 F 3.8
10:00 167.0 F 7.0
12:00 188.4 F 10.6
14:00 207.8 F 4.8
16:00 216.5 F 6.0
18:00 227.6 F 9.7
20:00 235.5 F 11.7
22:00 236.7 F 9.8
24:00 238.9 F 9.4
26:00 245.7 F 13.1
28:00 256.7 F 18.9
30:00 267.4 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 29.0
2:00 271.7 F 28.1
3:00 272.2 F 28.9
4:00 272.2 F 28.6

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 219.3 F 0.0
10:00 227.7 F 0.0
15:00 231.6 F 0.0
20:00 226.9 F 0.0
25:00 224.9 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

02/02/2022 09:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____