
Midmark M11 - v1.0.5
Total Cycles: 2573

Name: RITTER2
Sterilizer ID: V2313228
06/27/2023 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 172.1 F 1.3
4:00 177.2 F 4.7
6:00 209.6 F 9.2
8:00 224.8 F 9.5
10:00 228.3 F 8.3
12:00 237.5 F 13.2
14:00 237.3 F 10.5
16:00 236.9 F 8.7
18:00 248.8 F 15.2
20:00 259.9 F 21.4
22:00 270.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.9
2:00 272.2 F 29.4
3:00 271.4 F 28.9
4:00 271.9 F 29.6

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 204.8 F 0.0
10:00 212.0 F 0.0
15:00 214.3 F 0.0
20:00 215.0 F 0.0
25:00 213.8 F 0.0
30:00 212.1 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:00
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

06/27/2023 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____