
Midmark M11 - v1.0.0
Total Cycles: 1030

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2021 06:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 77.4 F 0.0
4:00 107.8 F 0.0
6:00 138.5 F 0.1
8:00 157.7 F 0.1
10:00 175.1 F 0.2
12:00 193.0 F 1.6
14:00 207.3 F 4.6
16:00 219.2 F 7.6
18:00 228.1 F 11.2
20:00 231.9 F 6.4
22:00 239.3 F 11.2
24:00 236.5 F 8.4
26:00 242.0 F 10.9
28:00 246.7 F 13.3
30:00 258.4 F 19.7
32:00 269.1 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.7
2:00 271.5 F 27.9
3:00 272.7 F 28.8
4:00 271.8 F 28.0

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 197.6 F 0.0
10:00 212.0 F 0.0
15:00 226.6 F 0.0
20:00 232.2 F 0.0
25:00 223.3 F 0.0
30:00 181.1 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:08
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:11:07
CYCLE COMPLETE

07/20/2021 07:49 AM
ACCEPTED: X
REJECTED:
Operator: U01