
Midmark M11 - v1.0.0
Total Cycles: 3268

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.3 F 0.0
2:00 197.8 F 1.7
4:00 217.7 F 7.2
6:00 227.7 F 12.7
8:00 231.9 F 8.5
10:00 242.4 F 11.4
12:00 243.2 F 12.1
14:00 241.3 F 10.8
16:00 254.6 F 17.5
18:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.9
2:00 272.3 F 28.5
3:00 271.5 F 27.9
4:00 271.9 F 28.5

Min 271.4 F 27.9
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 223.2 F 0.0
10:00 231.6 F 0.0
15:00 227.3 F 0.0
20:00 224.5 F 0.0
25:00 222.9 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

05/23/2023 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____