
Midmark M11 - v1.0.0
Total Cycles: 2774

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 180.6 F 0.0
4:00 197.6 F 4.5
6:00 215.5 F 10.0
8:00 225.0 F 6.5
10:00 236.5 F 11.3
12:00 235.7 F 9.0
14:00 238.6 F 9.4
16:00 249.6 F 15.0
18:00 262.3 F 22.2
20:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.6 F 29.1
4:00 272.3 F 28.7

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 217.4 F 0.0
10:00 227.7 F 0.0
15:00 226.6 F 0.0
20:00 226.1 F 0.0
25:00 229.8 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

01/12/2023 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____