
Midmark M11 - v1.0.5
Total Cycles: 680

Name: RITTER2
Sterilizer ID: V2313228
01/11/2022 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 192.3 F 2.7
4:00 193.9 F 7.2
6:00 212.4 F 12.8
8:00 227.6 F 8.6
10:00 236.8 F 10.8
12:00 242.5 F 12.9
14:00 243.5 F 12.3
16:00 256.6 F 19.3
18:00 268.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.5 F 29.6
2:00 272.1 F 29.0
3:00 271.3 F 28.9
4:00 272.6 F 29.7

Min 271.2 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 224.4 F 0.0
10:00 224.4 F 0.0
15:00 220.9 F 0.0
20:00 218.0 F 0.0
25:00 227.8 F 0.0
30:00 229.7 F 0.0

Drying Complete

FILLING: 3:29
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

01/11/2022 11:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____