
Midmark M11 - v1.0.0
Total Cycles: 5373

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2025 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 70.9 F 0.4
4:00 85.5 F 2.0
6:00 121.6 F 4.3
8:00 151.3 F 6.8
10:00 170.1 F 9.9
12:00 189.5 F 13.3
14:00 205.7 F 3.5
16:00 215.0 F 6.6
18:00 225.7 F 10.0
20:00 236.1 F 10.7
22:00 235.8 F 9.5
24:00 239.9 F 10.3
26:00 244.1 F 12.4
28:00 255.0 F 18.1
30:00 265.1 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.1 F 29.1
2:00 271.2 F 28.5
3:00 272.0 F 29.1
4:00 271.3 F 28.4

Min 271.1 F 28.4
Max 272.3 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 215.8 F 0.0
10:00 184.1 F 0.0
15:00 202.4 F 0.0
20:00 202.4 F 0.0
25:00 192.2 F 0.0
30:00 201.1 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:60
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:10:31
CYCLE COMPLETE

02/27/2025 10:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____