
Midmark M11 - v1.0.5
Total Cycles: 2159

Name: RITTER2
Sterilizer ID: V2313228
03/20/2023 07:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 84.5 F 0.0
4:00 117.8 F 0.0
6:00 130.8 F 0.0
8:00 161.5 F 1.5
10:00 182.8 F 5.1
12:00 204.6 F 9.0
14:00 215.7 F 12.9
16:00 221.3 F 7.1
18:00 233.2 F 11.4
20:00 231.6 F 7.7
22:00 243.0 F 13.3
24:00 242.4 F 11.8
26:00 254.3 F 18.0
28:00 265.2 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.8 F 29.5
2:00 272.6 F 29.2
3:00 272.2 F 29.2
4:00 271.8 F 29.3

Min 270.9 F 28.7
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 227.4 F 0.0
10:00 239.9 F 0.0
15:00 182.9 F 0.0
20:00 178.8 F 0.0
25:00 178.0 F 0.0
30:00 178.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:43
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:07:43
CYCLE COMPLETE

03/20/2023 08:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____