
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000020
13:15 08/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:16
STERILIZING 3:14
136.0 C 220kPa 3:14
136.2 C 224kPa 3:44
136.4 C 221kPa 4:14
136.4 C 225kPa 4:44
136.4 C 225kPa 5:14
136.4 C 225kPa 5:44
136.2 C 221kPa 6:14
136.4 C 224kPa 6:44
136.4 C 224kPa 6:44
VENTING 6:44
AIR DRYING 7:34
CYCLE COMPLETE 37:34

Digital Signature #
3D131AF66323C254
