
Midmark M11 - v1.0.5
Total Cycles: 2524

Name: RITTER2
Sterilizer ID: V2313228
06/15/2023 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 184.2 F 3.8
4:00 198.9 F 8.9
6:00 220.5 F 8.8
8:00 231.1 F 10.4
10:00 233.5 F 7.8
12:00 241.2 F 11.2
14:00 245.9 F 13.8
16:00 258.5 F 20.6
18:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.7 F 29.9
2:00 271.8 F 29.2
3:00 272.0 F 29.8
4:00 272.6 F 29.8

Min 271.3 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 216.2 F 0.0
10:00 221.5 F 0.0
15:00 224.9 F 0.0
20:00 226.1 F 0.0
25:00 226.8 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

06/15/2023 03:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____