

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1494

Name: RITTER2  
Sterilizer ID: V2313228  
09/15/2022 07:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.8 F 0.0  
2:00 169.4 F 1.9  
4:00 180.8 F 5.8  
6:00 210.3 F 11.1  
8:00 221.4 F 6.7  
10:00 235.5 F 12.3  
12:00 236.2 F 9.9  
14:00 237.5 F 8.7  
16:00 248.5 F 15.1  
18:00 260.6 F 21.9  
20:00 271.8 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.6 F 29.6  
2:00 271.5 F 28.9  
3:00 272.6 F 29.7  
4:00 271.9 F 29.1

Min 271.4 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 203.2 F 0.0  
10:00 219.2 F 0.0  
15:00 225.3 F 0.0  
20:00 227.7 F 0.0  
25:00 225.2 F 0.0  
30:00 222.6 F 0.0  
Drying Complete

FILLING: 2:35  
HEATING: 20:34  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:00:15  
CYCLE COMPLETE

09/15/2022 08:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_