
Midmark M11 - v1.0.0
Total Cycles: 3762

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2023 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 158.2 F 0.0
4:00 180.1 F 2.7
6:00 201.8 F 6.0
8:00 217.8 F 10.5
10:00 223.7 F 6.4
12:00 232.7 F 10.4
14:00 233.9 F 8.4
16:00 244.7 F 13.3
18:00 243.8 F 12.1
20:00 255.7 F 18.3
22:00 266.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.8
2:00 271.3 F 28.1
3:00 272.5 F 29.0
4:00 271.9 F 28.4

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 212.6 F 0.0
10:00 231.1 F 0.0
15:00 231.2 F 0.0
20:00 228.2 F 0.0
25:00 228.8 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:39
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:16
CYCLE COMPLETE

09/12/2023 04:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____