
Midmark M11 - v1.0.5
Total Cycles: 4414

Name: RITTER2
Sterilizer ID: V2313228
03/07/2025 05:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 155.7 F 2.7
4:00 182.8 F 7.2
6:00 206.7 F 13.1
8:00 225.3 F 7.6
10:00 237.7 F 13.0
12:00 237.6 F 10.7
14:00 236.7 F 9.4
16:00 251.5 F 16.7
18:00 263.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.7 F 29.6
2:00 272.1 F 29.6
3:00 272.4 F 29.3
4:00 272.3 F 29.7

Min 271.6 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.9 F 0.0
5:00 199.1 F 0.0
10:00 206.4 F 0.0
15:00 215.9 F 0.0
20:00 233.0 F 0.0
25:00 233.4 F 0.0
30:00 227.4 F 0.0

Drying Complete

FILLING: 1:31
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

03/07/2025 06:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____