
Midmark M11 - v1.0.0
Total Cycles: 4874

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2024 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.1 F 0.0
2:00 119.5 F 0.1
4:00 147.2 F 0.4
6:00 177.7 F 0.9
8:00 200.4 F 4.0
10:00 215.9 F 7.6
12:00 224.4 F 11.7
14:00 227.5 F 6.2
16:00 239.1 F 11.1
18:00 234.8 F 8.2
20:00 242.6 F 11.0
22:00 246.9 F 13.4
24:00 258.9 F 19.9
26:00 269.8 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.6 F 28.6
2:00 271.6 F 28.0
3:00 272.7 F 28.7
4:00 271.8 F 27.9

Min 271.5 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 233.3 F 0.0
10:00 221.7 F 0.0
15:00 211.5 F 0.0
20:00 227.3 F 0.0
25:00 231.2 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:57
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:05:19
CYCLE COMPLETE

09/06/2024 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____