

-----  
Midmark M11 - v1.0.5  
Total Cycles: 683

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2022 02:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
2:00 173.1 F 3.2  
4:00 184.6 F 7.3  
6:00 205.0 F 12.5  
8:00 221.7 F 7.4  
10:00 235.1 F 13.0  
12:00 237.5 F 11.0  
14:00 238.2 F 10.0  
16:00 252.1 F 17.0  
18:00 264.0 F 24.1  
20:00 271.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 271.8 F 29.6  
2:00 272.7 F 30.0  
3:00 272.8 F 29.8  
4:00 272.5 F 29.6  
  
Min 271.1 F 28.8  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 218.5 F 0.0  
10:00 236.5 F 0.0  
15:00 228.8 F 0.0  
20:00 223.2 F 0.0  
25:00 220.0 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 2:60  
HEATING: 20:07  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:00:12  
CYCLE COMPLETE

01/11/2022 03:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_