
Midmark M11 - v1.0.0
Total Cycles: 2254

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 07:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 75.1 F 0.0
4:00 97.3 F 0.0
6:00 106.2 F 0.1
8:00 141.1 F 0.4
10:00 170.7 F 0.7
12:00 191.5 F 1.3
14:00 209.7 F 3.3
16:00 222.3 F 6.5
18:00 231.2 F 9.6
20:00 239.7 F 13.3
22:00 235.8 F 9.4
24:00 240.7 F 10.5
26:00 244.8 F 12.6
28:00 242.1 F 11.1
30:00 254.6 F 17.4
32:00 265.1 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.6 F 28.7
2:00 271.6 F 27.9
3:00 272.2 F 28.6
4:00 272.2 F 28.3

Min 271.3 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.9 F 0.0
5:00 202.9 F 0.0
10:00 221.0 F 0.0
15:00 220.8 F 0.0
20:00 236.3 F 0.0
25:00 234.3 F 0.0
30:00 226.2 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 33:58
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:11:60
CYCLE COMPLETE

08/29/2022 08:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____