

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001640  
12:12 30/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:30  
PRESSURIZING 1:35  
STERILIZING 2:33  
136.0 C 220kPa 2:33  
136.4 C 225kPa 3:03  
136.2 C 223kPa 3:33  
136.2 C 224kPa 4:03  
136.3 C 225kPa 4:33  
136.4 C 225kPa 5:03  
136.4 C 225kPa 5:33  
136.4 C 226kPa 6:03  
136.4 C 226kPa 6:03  
VENTING 6:03  
AIR DRYING 6:55  
CYCLE COMPLETE 6:59  
-----

Digital Signature #  
046FFA380E8A2B7A  
-----