
Midmark M11 - v1.0.0
Total Cycles: 1694

Name: RITTER 1
Sterilizer ID: V2178198
03/02/2022 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 189.0 F 3.8
4:00 192.5 F 8.5
6:00 216.8 F 10.9
8:00 227.9 F 8.0
10:00 239.3 F 13.3
12:00 238.4 F 10.7
14:00 237.5 F 9.1
16:00 251.1 F 16.1
18:00 263.3 F 23.0
20:00 271.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.2
2:00 271.6 F 28.5
3:00 271.9 F 29.0
4:00 272.3 F 29.0

Min 271.2 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 231.1 F 0.0
10:00 229.2 F 0.0
15:00 228.0 F 0.0
20:00 230.6 F 0.0
25:00 228.6 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

03/02/2022 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____