
Midmark M11 - v1.0.5
Total Cycles: 1962

Name: RITTER2
Sterilizer ID: V2313228
01/24/2023 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.9 F 0.0
2:00 89.0 F 0.0
4:00 97.9 F 0.0
6:00 114.0 F 0.0
8:00 145.5 F 0.7
10:00 177.9 F 1.7
12:00 199.6 F 4.9
14:00 214.0 F 8.8
16:00 225.3 F 12.8
18:00 230.1 F 8.1
20:00 240.7 F 12.8
22:00 238.9 F 10.5
24:00 235.6 F 8.7
26:00 249.4 F 15.2
28:00 260.6 F 21.5
30:00 270.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.5
2:00 271.8 F 28.7
3:00 272.0 F 29.4
4:00 272.5 F 29.3

Min 271.2 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 220.4 F 0.0
10:00 199.4 F 0.0
15:00 195.5 F 0.0
20:00 195.9 F 0.0
25:00 198.4 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

01/24/2023 09:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____