
Midmark M11 - v1.0.5
Total Cycles: 1577

Name: RITTER2
Sterilizer ID: V2313228
10/06/2022 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.7 F 0.0
2:00 205.2 F 0.6
4:00 215.5 F 5.8
6:00 223.1 F 11.5
8:00 229.2 F 7.9
10:00 242.1 F 13.0
12:00 242.4 F 12.4
14:00 242.4 F 11.9
16:00 255.9 F 19.0
18:00 267.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.2 F 29.6
2:00 272.3 F 29.3
3:00 271.5 F 28.9
4:00 272.6 F 29.8

Min 271.4 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 217.3 F 0.0
10:00 232.7 F 0.0
15:00 229.9 F 0.0
20:00 224.4 F 0.0
25:00 222.9 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

10/06/2022 03:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____