
Midmark M11 - v1.0.0
Total Cycles: 2125

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 07:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 75.9 F 0.0
4:00 84.7 F 0.0
6:00 118.9 F 0.2
8:00 149.4 F 1.2
10:00 173.2 F 4.4
12:00 188.7 F 8.0
14:00 203.9 F 11.5
16:00 220.6 F 3.7
18:00 225.7 F 7.8
20:00 235.7 F 11.9
22:00 234.5 F 8.2
24:00 244.9 F 12.1
26:00 243.7 F 12.0
28:00 255.4 F 18.0
30:00 265.8 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.1 F 28.7
2:00 271.9 F 28.7
3:00 271.7 F 28.3
4:00 272.5 F 29.1

Min 271.1 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 195.3 F 0.0
10:00 202.7 F 0.0
15:00 194.9 F 0.0
20:00 207.9 F 0.0
25:00 211.1 F 0.0
30:00 212.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:49
CYCLE COMPLETE

07/14/2022 08:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____