
Midmark M11 - v1.0.5
Total Cycles: 660

Name: RITTER2
Sterilizer ID: V2313228
12/21/2021 03:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.4 F 0.0
2:00 108.5 F 0.5
4:00 124.6 F 2.8
6:00 152.3 F 6.1
8:00 179.8 F 10.3
10:00 208.9 F 6.0
12:00 221.3 F 8.4
14:00 229.3 F 13.2
16:00 233.1 F 9.9
18:00 236.1 F 9.2
20:00 245.8 F 13.9
22:00 257.0 F 20.2
24:00 268.0 F 27.0
26:00 272.2 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.6 F 30.1
2:00 272.1 F 30.1
3:00 272.5 F 30.0
4:00 272.0 F 30.1

Min 271.7 F 29.6
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 228.8 F 0.0
10:00 222.8 F 0.0
15:00 215.5 F 0.0
20:00 223.0 F 0.0
25:00 230.4 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:04:33
CYCLE COMPLETE

12/21/2021 04:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____