
Midmark M11 - v1.0.5
Total Cycles: 2134

Name: RITTER2
Sterilizer ID: V2313228
03/09/2023 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.5 F 0.0
2:00 199.7 F 4.3
4:00 220.7 F 9.9
6:00 224.5 F 6.6
8:00 236.2 F 11.8
10:00 236.7 F 9.6
12:00 236.9 F 9.1
14:00 250.6 F 15.9
16:00 262.9 F 23.2
18:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.2 F 29.5
2:00 271.4 F 28.9
3:00 272.3 F 29.7
4:00 271.9 F 29.1

Min 271.3 F 28.8
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 221.3 F 0.0
10:00 245.6 F 0.0
15:00 237.9 F 0.0
20:00 227.3 F 0.0
25:00 225.1 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:10
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:23
CYCLE COMPLETE

03/09/2023 01:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____