
Midmark M11 - v1.0.5
Total Cycles: 2264

Name: RITTER2
Sterilizer ID: V2313228
04/19/2023 07:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.4 F 0.0
2:00 78.6 F 0.0
4:00 83.5 F 0.0
6:00 110.8 F 0.0
8:00 148.8 F 2.9
10:00 174.8 F 6.0
12:00 194.4 F 9.5
14:00 213.7 F 11.8
16:00 221.9 F 7.4
18:00 231.9 F 11.8
20:00 231.0 F 8.3
22:00 242.0 F 12.4
24:00 242.5 F 12.2
26:00 254.9 F 18.5
28:00 265.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.3 F 29.4
2:00 272.6 F 29.9
3:00 271.7 F 29.4
4:00 272.5 F 29.6

Min 271.6 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 232.1 F 0.0
10:00 233.4 F 0.0
15:00 191.7 F 0.0
20:00 193.7 F 0.0
25:00 200.7 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:43
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:07:42
CYCLE COMPLETE

04/19/2023 08:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____