

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000807  
11:18 23/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:37  
PRESSURIZING 7:46  
STERILIZING 9:33  
136.0 C 220kPa 9:33  
136.5 C 222kPa 10:03  
136.2 C 219kPa 10:33  
135.8 C 218kPa 11:03  
136.3 C 222kPa 11:33  
136.4 C 223kPa 12:03  
136.5 C 221kPa 12:33  
136.3 C 221kPa 13:03  
136.3 C 221kPa 13:03  
VENTING 13:03  
AIR DRYING 14:24  
CYCLE COMPLETE 37:31

-----  
Digital Signature #  
CFDA6F5435665707  
-----